

HEC Certification (Listed): UL 61010-1, CSA C22.2#61010-1-12

Certification Mark



Certified Options

HEC

Cooling Capacity
002
006

-

Radiating Method	Power Supply	Communication
A	5	A
		B

-

Option(s)
(nil)
F
N

AUTHORIZATION TO MARK

This authorizes the application of the Certification Mark(s) shown below to the models described in the Product(s) Covered section when made in accordance with the conditions set forth in the Certification Agreement and Listing Report. This authorization also applies to multiple listee model(s) identified on the correlation page of the Listing Report.

This document is the property of Intertek Testing Services and is not transferable. The certification mark(s) may be applied only at the location of the Party Authorized To Apply Mark.

Applicant: SMC Corp. of America

Manufacturer: SMC Corporation

Address: 14191 Myford Road, Tustin, CA 92780

Address: 55-2 Kubo Odagawa, Yamatsuri-machi, Higashishirakawa-gun, Fukushima

Country: USA

Country: Japan

Contact: David Phung

Contact: Yoshio Kanazawa

Phone: (714) 338-5821

Phone: 0247-46-3000

FAX: (714) 669-0617

FAX: 0198-69-3300

Email: dphung@smcusa.com

Email: YOSHIOK@SMCJPN.CO.JP

Party Authorized To Apply Mark: Same as Manufacturer
Report Issuing Office: Intertek Japan K.K.

Control Number: 3054524

Authorized by: _____

Thomas J. Patterson, Certification Manager



This document supersedes all previous Authorizations to Mark for the noted Report Number.

This Authorization to Mark is for the exclusive use of Intertek's Client and is provided pursuant to the Certification agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this Authorization to Mark. Only the Client is authorized to permit copying or distribution of this Authorization to Mark and then only in its entirety. Use of Intertek's Certification mark is restricted to the conditions laid out in the agreement and in this Authorization to Mark. Any further use of the Intertek name for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. Initial Factory Assessments and Follow up Services are for the purpose of assuring appropriate usage of the Certification mark in accordance with the agreement, they are not for the purposes of production quality control and do not relieve the Client of their obligations in this respect.

Intertek Testing Services NA Inc.
545 East Algonquin Road, Arlington Heights, IL 60005
Telephone 800-345-3851 or 847-439-5667 Fax 312-283-1672

Standard(s):	Standard for Safety for Medical Electrical Equipment, Part 1: General Requirements for Safety, (UL 60601-1, First Edition, Dated April 25, 2003, with revisions through and including April 26, 2006) Medical Electrical Equipment, Part 1: General Requirements for Safety, (CAN/CSA-C22.2 NO. 601.1-M90, Reaffirmed November 2005)
Product:	Thermo-con Air-to-Liquid conditioner
Models:	HEC002-A5X-Y (where first X=A or B; second Y=F, N, or FN or X101)